

510(k) Summary of Safety and Effectiveness in accordance with 21 CFR 807.92

(a) (1) **Submitted by:** EnviteC-Wismar GmbH by Honeywell
Alter Holzhafen 18
Wismar 23966
Germany
Tel.: +49-38 41 360 221
Fax: +49-38 41 360 222
Marcus.Lindenlaub@Honeywell.com

Contact Person: Marcus Lindenlaub

Position/Title: Managing Director

Date of Preparation: June 10, 2009

(2) **Trade Name:** EnviteC Reusable SpO₂ Sensors

Common/Classification Name: OXIMETER; EAR OXIMETER

Product Code(s): DQA; 21 CFR §870.2700
DPZ; 21 CFR §870.2710

Class: Class II

(3) **Predicate Device(s):** Substantial Equivalence to:

K Number	Model	Manufacturer
K062605	Philips Reusable SpO ₂ Sensors, Models M1196A and M1196T (Finger Clip Sensors)	Philips Medical Systems (formerly HP)
K030973	Philips picoSAT II SpO ₂ Pulse Oximetry Module	Philips Medical Systems (formerly HP)
K053419	PULSOX-300, 300i	Konica Minolta Sensing, Inc.

Reason for Submission: New Device(s)

(4) **Description of Device:**

The EnviteC Reusable SpO₂ Sensors are a family of oximeter sensors designed and validated for compatibility with the predicate oximeter manufacturers listed above.

EnviteC's Reusable SpO₂ Sensors are comprised of a connector and a cable which terminates into a sensor housing. One housing half contains a dual LED light source, and the other half contains a light sensitive photodetector for pulse oximetry by transmittance method. Three types of sensor housings are offered in this submission:

- A finger clip with rigid halves positioned by mild spring force
- An ear clip with rigid halves positioned by mild spring force and an ear hanger for the cable
- A soft rubber finger sensor with unitary sealed tube type construction

A unique sensor type exists for each compatible manufacturer above, and each sensor has unique labeling and specifications designed for compatibility with the specific monitor manufacturer (HP/Philips, Minolta).

Each sensor type includes the following features:

- Connector pin-outs specific for the manufacturer type
- Component specifications specific for the manufacture type

Each sensor clearly specifies the manufacturer type with two compatibility statements:

- One printed on or attached to the sensor
- One on the instructions for use.

(5) Intended use:

The measurement of functional oxygen saturation of arterial hemoglobin (SpO₂) has been a standard of care in the USA for 20 years. Applications for oximetry include monitoring in the anesthesia, recovery, and critical care environments, as well as transport monitoring and home care.

Indications for Use:

EnviteC Reusable SpO₂ Sensors are indicated for continuous non-invasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO₂) and pulse rate for adult and pediatric (excluding neonatal and infant) patients in hospitals, hospital-type facilities, mobile units and home environments.

Prescription device.

(6) Technological Characteristics:

The EnviteC Reusable SpO₂ Sensors employ the same technological characteristics as the predicate devices to determine arterial oxygen saturation: arterially perfused tissue is illuminated sequentially by two wavelengths of light emitted by light emitting diodes (LED's), and the time

varying absorbance of the tissue is measured from a photodiode light sensor. This method is characteristic of all pulse oximeter sensors which are the subject of this submission as well as the predicate devices.

(b) (1) **Non-Clinical Tests Submitted:**

The sensors were tested in accordance with current applicable standards for medical device Electrical Safety and Electromagnetic Compatibility. The sensors (with predicate device monitors) were tested for pulse rate with a listed simulator. Sensor electro-optical parameters were measured and compared to predicate devices. The sensors passed all of the tests.

Sensor patient contact materials meet applicable standards for biocompatibility.

(2) **Clinical Tests Submitted:**

Clinical testing was performed to validate the performance and accuracy of the EnviteC Reusable SpO₂ Sensors under controlled hypoxia versus arterial oxygen saturation as determined by co-oximetry. All testing was performed under an institutionally approved protocol with subject informed consent. Clinical test results support the stated accuracy claims for the specified range of 70% to 100% SaO₂.

(3) **Conclusions from Tests:**

As described in (b)(1) and (b)(2) above, EnviteC Reusable SpO₂ Sensors are equivalent to predicate sensors as substantiated by parameter, bench, and clinical testing. Device safety is substantiated by testing to applicable standards and by biocompatibility of patient contact materials.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

EnviteC-Wismar GmbH
C/O Mr. Stephen H. Gorski
Imagenix, Incorporated
S65 W35739 Piper Road
Eagle, Wisconsin 53119

JUL 24 2009

Re: K090438

Trade/Device Name: EnviteC Reusable SpO₂ Sensors
Regulation Number: 21 CFR 870.2700
Regulation Name: Oximeter
Regulatory Class: II
Product Code: DQA, DPZ
Dated: June 23, 2009
Received: June 24, 2009

Dear Mr. Gorski:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Susan Runner, D.D.S., M.A.

Acting Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use**510(k) Number (if known):** K090438**Device Name:** EnviteC Reusable SpO₂ Sensors**Indications for use:**

EnviteC Reusable SpO₂ Sensors are indicated for continuous non-invasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO₂) and pulse rate for adult and pediatric (excluding neonatal and infant) patients in hospitals, hospital-type facilities, mobile units and home environments.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND / OR

Over-The-Counter Use
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K090438Page 1 of